



Declaration of Conformity

Illumina, Inc. hereby declares under its sole responsibility that the product(s) listed are in conformity to the EMC Directive [2004/108/EC], Low Voltage Directive [2006/95/EC].

MANUFACTURER:	Illumina, Inc	FACTORY LOCATION:
ADDRESS:	5200 Illumina Way San Diego, CA 92122, USA	Illumina Singapore Pte. Ltd North Tech Lobby 3 #02-13118 29 Woodlands Industrial Park E1 Singapore, 757716
PRODUCT TYPE:	Genetic Sequencer	
MODEL:	MiniSeq	
CE MARK AFFIXED:	2016	

The construction of the product is in compliance with the following harmonized and/or consensus standards.

IEC 61010-1:2010 UL 61010-1:2012 IEC/EN 61010-1:2010 CAN/CSA C22.2 No. 61010-1-12	<i>Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements</i>
IEC 61010-2-081:2015 UL 61010-2-81:2015 CAN/CSA C22.2 No. 61010-2-081:15	<i>Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes</i>
IEC 60825-1:2014	<i>Safety of laser products – Part 1: Equipment classification and requirements</i>
IEC 61326-1:2012 (Class A)	<i>Electrical equipment for the measurement, control and Laboratory use – EMC Requirements Part1, Class A</i>
EN 61326-2-6:2012	<i>EMC requirements -- Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment</i>
EN 50364:2001	<i>Limitation of human exposure to electromagnetic fields from devices operating in the frequency range 0 Hz to 300 GHz, used in electronic article surveillance (EAS), radio frequency identification (RFID) and similar applications</i>
ETSI EN 300 330-1 V1.8.1	<i>Electromagnetic compatibility and Radio spectrum Matters (ERM); Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz; Part 1: Technical characteristics and test methods</i>
ETSI EN 300 330-2 V1.6.1	<i>Part 2: Harmonized EN covering the essential requirements of article 3.2 of the R&TTE Directive</i>
ETSI EN 301 489-3 V1.5.1	<i>Electromagnetic compatibility and Radio spectrum Matters (ERM); Electro Magnetic Compatibility (EMC) standard for radio equipment and services; Part 3: Specific conditions for Short-Range Devices (SRD) operating on frequencies between 9 kHz and 246 GHz</i>
Title 47, CFR Part 15 Subpart C § 15.225	<i>Intentional radiators operating within the band 13.110–14.010 MHz.</i>
Title 47, CFR Part 15 Subpart B	<i>Unintentional radio frequency devices</i>

Authorized by:


 Mya Thomae
 VP, Regulatory and Clinical Affairs

1-19-16
 Date